



PROFICIENCY TESTING PLAN

ZZ 2026/1 – Soil Testing

Poskytovatel zkoušení způsobilosti při SZK FAST

Veveří 95, 602 00 Brno

Czech Republic

www.szk.fce.vutbr.cz

www.ptprovider.cz

Coordinator

Supervisor

Approved for PT
Provider

Assoc. Prof. Ing. Tomáš Vymazal, Ph.D.

Ing. Petr Misák, Ph.D.

Assoc. Prof. Ing. Tomáš Vymazal, Ph.D.

Approved: August 21, 2026

Approved: August 21, 2026

Approved: August 21, 2026

Contents

1 Basic Information about the Proficiency Testing Program	2
2 Implementation of the Proficiency Testing Program	2
2.1 Specifications and Characteristics	2
2.2 Ensuring Homogeneity and Stability	6
2.3 Instructions for Eliminating Major Sources of Errors and Risks	6
2.4 PTP Schedule	7
3 Procedures used in the Statistical Analysis of Laboratory Results	7
4 Certificate of Participation and the Final Report on the Results of Interlaboratory Comparison	7
5 Safeguards for Confidentiality	7
6 Related Documents	7

1 Basic Information about the Proficiency Testing Program

The aim of the Proficiency Testing Program (PTP) is to compare and evaluate the results of laboratory tests conducted on soils in compliance with selected parts of ČSN EN ISO 17892-1 [1], ČSN EN ISO 17892-3 [2], ČSN EN ISO 17892-4, čl. 5.2 [3], ČSN EN ISO 17892-4, čl. 5.3 [3], ČSN EN ISO 17892-5 [4], ČSN EN ISO 17892-7 [5], ČSN EN ISO 17892-10 [6], ČSN EN ISO 17892-12, čl. 5.3 [7], ČSN EN ISO 17892-12, čl. 5.4 [7], ČSN EN 13286-2 [8] and ČSN EN 13286-47 [9].

The program strives to provide objective information about the measuring skills of PTP participants. The basic criterion for participation is timely registration for the program, and the prerequisites for obtaining the Certificate of Participation and the Final Report on the Results of Interlaboratory Comparison are timely payment of the fee and adherence to the schedule.

Important dates:

Registration deadline:	August 31, 2026
Distribution of samples:	October 12, 2026
Realization/initiation of testing:	November 9, 2026
Results sent to the organizer:	November 27, 2026
Presentation of the Certificate of Participation:	January 31, 2027

Submission of test results takes place exclusively via the MyPT customer zone at <https://test.ptprovider.cz/zona>. Results are entered by the laboratory after signing in to their account, in the section of the respective programme.

2 Implementation of the Proficiency Testing Program

2.1 Specifications and Characteristics

Testing laboratories and other institutions interested can register for the PTP. The minimum number of participants is 5. If the number of participants is close to the minimum, the coordinator will consider the evaluation of PTP results using Horn's procedure to determine the assigned value and measurement uncertainty. The maximum number of participants is 30. If the minimum number of participants is not reached, the PT Provider reserves the right to cancel the PTP. This takes place according to Chapter 3 of the "Cancellation and Complaint Proceedings" instructions [10] available on www.ptprovider.cz.

Parts of the PT program:

ZZ01 ČSN EN ISO 17892-1 [1]

- Characteristics: Water content
- Units: %
- Range of observed parameters: 1 – 20 %
- Number of determinations: 3

ZZ02 ČSN EN ISO 17892-3 [2]

- Characteristics: Particle density
- Units: Mg/m³
- Range of observed parameters: 2,6 – 2,75 Mg/m³
- Number of determinations: 3

ZZ03 ČSN EN ISO 17892-4, čl. 5.2 [3]

- Characteristics: Particle size distribution
- Units: %
- Range of observed parameters: It can not be determined in advance
- Material: unwashed aggregate 0/4 mm, approx. 0,5 kg
- Expected sieves used: 8, 4, 2, 1, 0.5, 0.25, 0.125, 0.063 mm
- Number of determinations: 1

ZZ04 ČSN EN ISO 17892-4, čl. 5.3 [3]

- Characteristics: Values of weight proportions
- Units: %
- Range of observed parameters: It can not be determined in advance
- Material: loess, approx. 0.5 kg
- Equivalent particle diameter: 0,002, 0,005, 0,010, 0,020, 0,050 mm
- Number of determinations: 1
- **Instructions:**
 1. Read the values of the weight fractions of fractions smaller than the above equivalent particle diameter directly from the grain size curve or calculate them by interpolation between the nearest measured points on the grain size curve. Report the result in % rounded to 1 decimal place.
 2. Use the apparent particle density value of 2.65 Mg/m³ in your calculations.

ZZ05 ČSN EN ISO 17892-5 [4]

- Characteristics: Incremental loading oedometer test
- Units: MPa
- Range of observed parameters: It can not be determined in advance
- Number of determinations: 1
- **Instructions:**
 1. Dry the required amount of soil at 105°C.
 2. Homogenize the soil.
 3. Sieve through a mesh with 4 mm aperture.
 4. Moisturize to the prescribed water content 15.5 %.

5. Leave to rest in a plastic bag for 24 hours.
6. Compact with Proctor Standard energy.
7. Cut out samples for the indicated tests from the compacted soil.
8. Cut out the sample from approximately half the height of the compacted soil.
9. Pour water over the sample after loading to 50 kPa. (consolidation level). This step will not be evaluated under the PT scheme.
10. Perform the test at stress levels: 100, 200 and 400 kPa
11. No reconsolidation.
12. Evaluate the secant modulus values between stress levels 50–100, 100–200, 200–400 kPa - see standard EN ISO 17892-5 [4].
13. Indicate the moisture and dry bulk density achieved before the test.

ZZ06 ČSN EN ISO 17892-7 [5]

- Characteristics: Unconfined compressive strength, Strain at failure
- Units: MPa, %
- Range of observed parameters: It can not be determined in advance
- Number of determinations: 4
- **Instructions:**
 1. Dry the required amount of soil at 105°C.
 2. Homogenize the soil.
 3. Sieve through a mesh with 4 mm aperture.
 4. Moisturize to the prescribed water content 15.5 %.
 5. Leave to rest in a plastic bag for 24 hours.
 6. Compact with Proctor Standard energy.
 7. Cut out samples for the indicated tests from the compacted soil.
 8. Perform the test with two cylinder specimens of 38 mm in diameter and 76 mm in height.
 9. Compress the cylinders at the rate of 1 mm/min.
 10. Calculate the average value.
 11. Indicate the moisture and dry bulk density achieved before the test.

ZZ07 ČSN EN ISO 17892-10 [6]

- Characteristics: Shear stress at the vertical stress levels 50, 100, 200 and 400 kPa
- Units: kPa
- Range of observed parameters: It can not be determined in advance
- Number of determinations: 1
- **Instructions:**
 1. Dry the required amount of soil at 105°C.
 2. Homogenize the soil.

3. Sieve through a mesh with 4 mm aperture.
4. Moisturize to the prescribed water content 15.5 %.
5. Leave to rest in a plastic bag for 24 hours.
6. Compact with Proctor Standard energy.
7. Cut out samples for the indicated tests from the compacted soil.
8. Perform the test with four specimens at vertical stress levels: 50, 100, 200 and 400 kPa.
9. After loading, pour water over the samples and leave to consolidate for 24 hours.
10. Shear the samples at the rate 0,01 mm/min.
11. Indicate the moisture and dry bulk density achieved before the test.

ZZ08 ČSN EN ISO 17892-12, čl. 5.3 [7]

- Characteristics: Atterberg limits (Cone Penetration Method)
- Units: -
- Range of observed parameters: 20 – 70, 10 – 30
- Number of determinations: 3
- **Instructions:** The measured quantity depends on the type of soil.
- Determination of liquid limit:
 - Instrumentation according to the article 4.2 [7]
 - Procedure according to the article 5.3 [7]
 - Calculation according to the article 6.2 [7]
- Determination of plasticity limit:
 - Instrumentation according to the article 4.4 [7]
 - Procedure according to article 5.5 [7]
 - Calculation according to article 6.4 [7]

ZZ09 ČSN EN ISO 17892-12, čl. 5.4 [7]

- Characteristics: Determination of Liquid Limit (Casagrande)
- Units: -
- Range of observed parameters: 20 – 70
- Number of determinations: 3
- **Instructions:**
 - The measured quantity depends on the type of soil.
 - Instrumentation according to the article 4.3 [7]
 - Procedure according to the article 5.4 [7]
 - Calculation according to the article 6.3 [7]

ZZ10 ČSN EN 13286-2 [8]

- Characteristics: Proctor Standard

- Units: kg/m³, %
- Range of observed parameters: It can not be determined in advance.
- Number of determinations: 1 (from 5 points)
- **Instructions:** Use a A type Proctor compaction mold.

ZZ11 ČSN EN 13286-47 [9]

- Characteristics: IBI (test without saturation)
- Units: %
- Range of observed parameters: It can not be determined in advance
- Number of determinations: 1
- **Instructions:**
 1. Moisturise to the prescribed water content 16 % and compact with Proctor Standard energy,
 2. Do not let the compacted sample mature in water, test it immediately,
 3. Indicate the moisture and dry bulk density achieved before the test.

2.2 Ensuring Homogeneity and Stability

PT Provider employees and any suppliers they may utilize are aware of the significance of the homogeneity and stability of test specimens for the results of the Proficiency Testing Program. The homogeneity and stability of specimens is ensured in the following ways:

1. preparing the material for the preparation of samples from one bearing,
2. the distribution of bodies made of more bearings so as to ensure homogeneity of bodies in the field of testing of related characteristics,
3. by review the material before releasing participants.

2.3 Instructions for Eliminating Major Sources of Errors and Risks

PTP participants have the obligation:

- to handle the proficiency testing materials in the same way they handle the majority of routinely tested samples,
- to follow the instructions of the PT Provider employee responsible for the PTP, especially regarding the type of testing carried out, the number of result determinations and the PT schedule,
- to state measurement uncertainties in accordance with their documented procedures, including the corresponding expansion coefficient. Participants will use expansion coefficient 2, which approximately represents the 95 % reliability level, unless stated otherwise,
- Adhere to the rules and principles of ethical behavior, avoiding unfair practices that could negatively impact the evaluation of the PT program,
- adhere to the rules and principles of ethical behavior, avoiding unfair practices that could negatively impact the evaluation of the PT program,

- follow occupational health and safety and fire protection regulations, using only electrical equipment and instruments with valid inspections,
- to send the test results obtained during proficiency testing, including measurement uncertainties, to the PT Provider by the set deadline (see part 1).

2.4 PTP Schedule

All other information, forms and records not included in this document are accessible in updated form at <http://ptprovider.cz/?lang=en>.

3 Procedures used in the Statistical Analysis of Laboratory Results

Procedures used in the statistical analysis of proficiency testing programs can be found at ptprovider.cz.

4 Certificate of Participation and the Final Report on the Results of Interlaboratory Comparison

The PT Provider gives expert commentary on participant efficiency evaluation in the Final Report as part of training courses the PT Provider organises. The Final Report preserves the anonymity of the PTP participants. Each participant, or the participant's test results, is represented by an ID number. The Certificate of Participation in the PT programme is part of the Final Report. The Certificate is unique to each participant and includes the participant's ID number.

5 Safeguards for Confidentiality

The identity of PTP participants is confidential and only known to persons/subjects involved with the PTP. All participant information is considered confidential by the PT Provider. The participant may renounce this confidentiality for the purposes of discussion and mutual assistance until the PTP results are obtained. The PT Provider reveals the proficiency testing results to no third party with the sole exception of a written request by a regulatory authority submitted prior to the commencement of the PTP and which has been granted a written consent by the PTP participants.

6 Related Documents

- Quality Handbook of the PT Provider at the SZK FAST
- Cancellation and Complaint Proceedings available at <http://ptprovider.cz/?lang=en> [10]
- MPA 20 – 01 - . . . for application of EN ISO/IEC 17043 Concordance Assessment – General Requirements for Proficiency Testing in the Accreditation System of the Czech Republic.

References

- [1] EN ISO 17892-1. *Geotechnical investigation and testing - Laboratory testing of soil - Part 1: Determination of water content*. 2015.
- [2] EN ISO 17892-3. *Geotechnical investigation and testing - Laboratory testing of soil - Part 3: Determination of particle density*. 2016.
- [3] EN ISO 17892-4. *Geotechnical investigation and testing - Laboratory testing of soil - Part 4: Determination of particle size distribution*. 2017.
- [4] EN ISO 17892-5. *Geotechnical investigation and testing - Laboratory testing of soil - Part 5: Incremental loading oedometer test*. 2017.
- [5] EN ISO 17892-7. *Geotechnical investigation and testing - Laboratory testing of soil - Part 7: Unconfined compression test*. 2018.
- [6] EN ISO 17892-10. *Geotechnical investigation and testing - Laboratory testing of soil - Part 10: Direct shear tests*. 2018.
- [7] EN ISO 17892-12. *Geotechnical investigation and testing - Laboratory testing of soil - Part 12: Determination of liquid and plastic limits*. 2018.
- [8] EN 13286-2. *Unbound and hydraulically bound mixtures - Part 2: Test methods for laboratory reference density and water content - Proctor compaction*. 2011.
- [9] EN 13286-47. *Unbound and hydraulically bound mixtures - Part 47: Test method for the determination of California Bearing ratio, immediate bearing index and linear swelling*. 2021.
- [10] *Cancellation and Complaint Proceedings* – available at www.ptprovider.cz.